



Certificate of Analysis

Product Description

Product Name	PriGrow III Medium
Cat Number	TM003
Lot Number	10013688
Expiration Date	09/2016
Storage	2-8°C

Quality Control

Parameters	Specifications	Results
pH	7.2 ± 0.2	7.2
Osmolality	335 ± 30 mOsm/Kg H ₂ O	341 mOsm/Kg H ₂ O
Endotoxin	≤ 0.25 EU/mL @ FW	<0.0010 EU/mL @ FW
Mycoplasma	Negative	Negative
Sterility	Pass	Pass

This product is for research use only and is not intended for therapeutic or diagnostic applications. Please contact a technical service representative for more information.

No. 8, 13520 Crestwood Place
Richmond BC, Canada V6V 2G2
Tel: 604-247-2416
Fax: 604-247-2414
www.abmGood.com